

Technical Services Product Bulletin No: 001

Global Marine Systems Limited;
A Member of the Universal Jointing Consortium
9th January 2003



www.ujuc.com

Type 13 X-Ray Cameras

Global Marine has updated the Operating Manuals for all marks of Type 13 X-Ray Cameras (OMM012 & OMM036) to include the following mandatory statement:

“Radiation levels above $2.5\mu\text{Sv/h}$ at any point more than 50mm from the exterior surface of the camera should be considered a fault condition and under no circumstances should the camera be used until it is repaired by a competent, trained person”.

This is to remove any ambiguity in the method used to measure radiation levels for the camera during routine maintenance.

